Artificial Eye Measurement Performed on 17-Apr-2014

Artificial eye measurements in three trials. The duration of each trial is 10 seconds.

	Trial 1	Trial 2	Trial 3	
Mean_Hor	0.07509	0.07516	0.07496	^
Mean_Ver	0.15420	0.15471	0.15598	=
Std_Hor	0.00126	0.00129	0.00126	
Std_Ver	0.00146	0.00152	0.00155	-

Figure 1: Mean and standard deviation in horizontal and vertical axis.




Figure 2: Top row represents the eye tracker output in horizontal and vertical axis for each trial. the bottom row are the PSD of the output calculated by Welch method. Each column refers to one trial.