



Monocular DSP Status (Right DPI)

March 10, 2018


Static Test



	Trial 1	Trial 2	Trial 3
Std_Hor(mv)	0.984	0.918	0.960
Std_Ver(mv)	0.943	0.956	0.957
Std_Hor(arcmin)	0.197	0.184	0.192
Std_Ver(arcmin)	0.189	0.191	0.191
rms_Hor(mv)	1.045	1.488	1.147
rms_Ver(mv)	1.309	1.097	1.130
rms_Hor(arcmin)	0.209	0.298	0.229
rms_Ver(arcmin)	0.262	0.219	0.226



Square Wave: 0.1V, 1Hz



Square Wave: 5V, 1Hz

